

Calibration Report for Amplitude Control Linearity

A Nondestructive testing (NDT) calibration report format suitable for Ultrasonic Equipment calibrations. This is a sample report and may be required to be modified as per specific requirements.

NDT Procedure No: TNE-DOC-UT-03 Rev '0'



TRINITY NDT®

An ISO 9001:2008 Certified Company

Plot No.V-22(A), 6th Main, 2nd Stage, Peenya Industrial Estate, Bangalore-560058, INDIA

Telefax:+91-80-28366466, Mobile:+91 98441 29439, 99009 29439,

E-mail: info@trinityndt.com, Website: www.trinityndt.com

This UT calibration report & other NDT reports are available for free download at
www.trinityndt.com



TRINITY NDT[®]
An ISO 9001:2008 Certified Company

AMPLITUDE CONTROL LINEARITY


Procedure No. TNE-UT-Doc-03 Rev '0.'

This procedure & other NDT procedures are available for free download at www.trinityndt.com

Instrument Model	:		SL.No:
------------------	---	--	--------

Manufacturer	:	
--------------	---	--

Applicable Code	:	
-----------------	---	--



Indication Set at % Of Full Screen	dB Control Change	Accepted % Of Full Screen	Indication % Of Full Screen
80 %	-6 dB	32 to 48 %	
80 %	-12 dB	16 to 24 %	
40 %	+6 dB	64 to 96 %	
20 %	+12 dB	64 to 96 %	

	Tested By	Approved By	Third Party(if required)
Name			
Sign.			
Level			
Date			



TRINITY NDT[®]

An ISO 9001:2008 Certified Company

Aerospace & General NDT | Welding Inspection | Training & Certification | ASNT Level III Consulting | Sales

Visit us: www.trinityndt.com