

Calibration Report Format for Amplitude Control Linearity

A Nondestructive testing (NDT) calibration report format suitable for Ultrasonic Equipment calibrations. This is a sample report and may be required to be modified as per specific requirements.



Trinity NDT WeldSolutions Private Limited

ISO17025 & NADCAP Accredited Labs | Centre of Excellence in Welding

491, Site No.12, 14th Cross, 4th Phase, Peenya Industrial Area, Bangalore-560058, INDIA

M: +91 98441 29439, 72040 29439, 91413 39969 | P: +91 80 41205245

E: info@trinityndt.com

*NDT procedures and sample reports formats are available for free download at
www.trinityndt.com*



TRINITY NDT WELDSOLUTIONS PRIVATE LIMITED

ISO17025 & NADCAP Accredited NDT Labs | Centre of Welding

ULTRASONIC TESTING(UT) AMPLITUDE CONTROL LINEARITY CALIBRATION FORMAT

Procedure No. TNE-UT-Doc-03 Rev '01'

This procedure & other NDT procedures are available for free download at www.trinityndt.com

Instrument Model	:	SL.No:
Manufacturer	:	
Applicable Code	:	
Sketch:		

Indication Set at % Of Full Screen	dB Control Change	Accepted % Of Full Screen	Indication % Of Full Screen
80 %	-6 dB	32 to 48 %	
80 %	-12 dB	16 to 24 %	
40 %	+6 dB	64 to 96 %	
20 %	+12 dB	64 to 96 %	

	Tested By	Approved By	Third Party(if required)
Name			
Sign.			
Level			
Date			



TRINITY NDT[®]

Aerospace NDT | Centre of Excellence in Welding | Training



www.trinityndt.com